



MINISTRY OF HEALTH OF MONGOLIA

Olympic street-2, Government building VIII,
Sukhbaatar district, Ulaanbaatar-14210, MONGOLIA,
Tel: (976-51) 26-38-22, Fax: (976-11) 32-35-41, 32-09-16
E-mail: moh@moh.gov.mn, http://www.moh.gov.mn

Date 21 December 2023
Ref. 2/307

TO: POTENTIAL BIDDERS

REFERENCE: Request for Quotations of Goods no. G14
Procurement of: Delivery and Installation of Computer
Tomography scanners – 3 sets and the Related Services
Mongolia Covid-19 Emergency Response and Health System
Preparedness Project

Dear Sir / Madam,

We have received a request for clarification from a potential bidder under the above-referenced procurement. The response to the request for clarification is as follows:

Question 1: We have one question about 'CT tech spec spreadsheet': As for item 106-specific certification required, our CT machine has been approved by NMPA (China FDA) which is a member of IMDRF, is NMPA acceptable to you?

Clarification 1: Certification issued by the National Medical Products Association (NMPA) of the People's Republic of China, a member of IMDRF, is acceptable.

Question 2: Clause 20.3 of Conditions of Contract specifies the warranty period to be 36-40 months, while Clause (f) of the Request for Quotation document specifies 24 months. Kindly clarify warranty period.

Clarification 2: The Warranty period is 24 months from the date of satisfactory installation as per Clause (f) of the Request for Quotation document, Clause 11 of Conditions of Contract, and item nos. 78, 100, 115, 130, 140, 151, 159 of 1.3.1 Technical Specifications and Price Schedule.

ACTING STATE SECRETARY

S.TUGSDELGER

144223743